



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

APPLIED TEST SYSTEMS
154 East Brook Lane
Butler, PA 16002
Devon Rottman Phone: 724 283 1212

CALIBRATION

Valid To: February 29, 2024

Certificate Number: 2132.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 5}:

I. Dimensional

Parameter/Equipment	Range	CMC ² (±)	Comments
Displacement ³ – Extensometers	(0.0002 to 0.140) in (0.141 to 0.500) in	0.000 071 in 0.000 092 in	ASTM E83 w/ micrometer heads
	(0.4 to 6) in (10 to 152) mm	0.0023 in 0.059 mm	Calipers and outside micrometers
	Up to 0.02 in (0.02 to 0.6) in (0.6 to 2) in	21 μin 54 μin 140 μin	Epsilon 3590 VHR



II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ² (±)	Comments
Electronic Calibration of Thermocouple Indicators –			
Type B	(600 to 1650) °C	0.58 °C	Fluke 724
Type J	(-150 to 1000) °C	0.58 °C	
Type S	(170 to 1500) °C	0.40 °C	
Type R	(170 to 1500) °C	0.54 °C	
Type N	(130 to 1300) °C	0.30 °C	
Type K	(-150 to 1300) °C	0.40 °C	

III. Mechanical

Parameter/Equipment	Range	CMC ^{2, 4, 6} (±)	Comments
Force ³ – Measure			
Tension and Compression	Up to 20.0 lbf Up to 50.0 lbf Up to 150 lbf	0.24 % 0.22 % 0.20 %	ASTM E4 w/ standard/ dead weights
Tension	Up to 10 000 lbf Up to 20 000 lbf Up to 60 000 lbf	0.059 % 0.062 % 0.061 %	ASTM E4 w/ proving rings
	(66.4 to 1500) lbf (816 to 12 000) lbf	0.086 % 0.051 %	ASTM E4 w/ S beam load cell
	(26 to 1000) lbf (200 to 10 000) lbf (800 to 25 000) lbf (5000 to 50 000) lbf	0.042 % 0.049 % 0.041 % 0.068 % FS	ASTM E4 w/ load cell

Parameter/Equipment	Range	CMC ^{2, 4, 6} (±)	Comments
Force ³ – Measure (cont)			
Compression	Up to 10 000 lbf Up to 20 000 lbf Up to 60 000 lbf Up to 300 000 lbf	0.059 % 0.075 % 0.098 % 0.23 %	ASTM E4 w/ proving rings
	(20 to 1000) lbf (200 to 10 000) lbf (796 to 25 000) lbf (5000 to 50 000) lbf	0.056 % 0.047 % 0.048 % 0.068 % FS	ASTM E4 w/ load cell
Vacuum, Pressure and Burst Systems ³ – Pneumatic	(15 to 100) kPa (2.00 to 2.15) MPa	0.16 kPa 0.0026 MPa	ASTM D6521, AASHTO R28
	(100 to 605) psi (605 to 3000) psi	0.41 psi 1.3 psi	ASME B40, ASTM D1599
Alignment ³ – Course	40 % or Less	1.7 %	ASTM E1012 w/ strain gage
Bending Beam Rheometer Gauge Kit ³ –			ASTM D6648 AASHTO T313
Special Weights	Up to 220 g	0.004 g	
Step Gage & Beam Thickness	(0.05 to 0.400) in (0 to 12) mm	0.000 19 in 0.0043 mm	Micrometer
Beam Widths & Lengths	(0.401 to 6) in (10.0 to 152) mm	0.0023 in 0.059 mm	Calipers

IV. Thermodynamics

Parameter/Equipment	Range	CMC ^{2, 6} (\pm)	Comments
Temperature ³ – Measure	(-40 to 180) °C	0.037 °C	Digital monitor w/ platinum RTD
Temperature – Measure	(15 to 30) °C (59 to 86) °F	0.35 °C 0.63 °F	Fluke 51 w/ thermocouple

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve w/in its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ This laboratory performs field calibration activities for these parameters. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, percentages are percentage of reading, unless otherwise indicated, FS is defined as full scale.

⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.

⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



Accredited Laboratory

A2LA has accredited

APPLIED TEST SYSTEMS

Butler, PA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 14TH day of February 2022.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2132.01
Valid to February 29, 2024

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.